

IMP H & J EIS
ELECTRONIC CALIBRATION

by

R. A. Mewaldt & S. B. Vidor

Table of Contents

I.	Introduction
II.	Electronic Calibration Procedure
III.	Results of Electronic Calibrations
IV.	Calibration of D2 and D5
V.	Problems in Interpreting Data
VI.	Miscellaneous Information
VII.	Appendices:
	A-H Selected IMP-H CALAN output
	A-J Selected IMP-H CALAN output
	B-H IMP-H Measured and Interpolated Channel Thresholds
	B-J IMP-J Measured and Interpolated Channel Thresholds

IMP-H and J-EIS Electronic Calibration

I. Introduction

The energy loss of charged particles in the solid state detectors aboard the IMP EIS instruments is converted into digitized information (channel #'s). The purpose of this report is to describe the calibration procedure used to determine the energy loss if the channel is known, to supply some calibration data, and to point out some of the problems in interpreting the data. The report is generally arranged in the chronological order of the calibration procedure.

II. Electronic Calibration Procedure

A. Initial Calibration

The first step in the calibration procedure is to obtain a calibrated standard. This was done with a standard breadboard test pulser, a standard breadboard test capacitor of 1/2 pf., a 100 μ thick solid state detector with an Am²⁴¹ alpha source mounted close to its active area, a precision voltage source, and an amplifier and pulse height analyzer system. Figure 1 illustrates the schematic setup. If the standard test pulser/test capacitor combination is linear and has no offset, then it is only necessary to determine the voltage from the precision voltage source that causes the same response in the pulse height analyzer as do the alpha particles in the solid state detector (see internal report #31).

The next step is to calibrate the internal test pulser-detector-ADC system in the instrument. This is done after the detector stack has been installed and connected to the ADC's. The setup is shown schematically in Figure 2. The voltage corresponding to each ADC threshold and to the thresholds of channels 20, 100, and 1000 are determined using the breadboard test pulser and the




Figure 1




Figure 2

instrument test pulser. (The threshold of channel n is obtained when 50% of the pulses are in channel n and 50% in channel $n-1$.) These points not only calibrate the internal test pulser against the breadboard test pulser but also confirm the linearity of the system. The final result of this calibration is a ratio R for each ADC internal test pulser:

$$R = E/V \quad (1)$$

where E is the energy loss in a detector and V is the voltage from the precision power supply. The estimated accuracy of each calibration step is 0.1%.

B. Normal Calibration

Because of small drifts in the electronics as a function of time or temperature (and to verify that the drifts are not large), the ADC's must be periodically calibrated with internal test pulsers. The procedure was automated with the help of a programmable voltage source and Curt Widdoes' Varian program called CALAN. Using an algorithm described below, the program calculates the thresholds for all ADC's (D0 through D11) and a certain set of individual channel thresholds for D0 through D9. The channels are chosen such that the thresholds of other channels can be interpolated with a high degree of accuracy (typically to better than 0.2 channels for the higher channels). For some reason, which is not understood, the CALAN program does not give a reliable result for the ADC threshold. (It does calibrate the channel thresholds reproducibly however.) Therefore, the initial ADC threshold calibration (Section II-A) is taken as the proper calibration.

The CALAN algorithm for calculating thresholds is now described. The 40% and 60% ADC threshold triggering voltages are obtained with either 1000 or 10000 pulses. Using these 2 points the program computes the ADC threshold and the standard deviation (σ). To calculate the threshold for channel n ,

the program finds a voltage between the 40% and 60% triggering points of channel n. It then uses the ADC threshold σ to calculate the 50% point for channel n. The program then multiplies the voltage by the ratio R (see equation 1) to obtain the threshold energy. The assumption that the ADC threshold σ equals the channel σ 's gives rise to an error in the threshold estimate of less than 1 Kev.

After calibrating individual channel thresholds, the program calculates three quantities: gain, offset, and linearity, which are defined below. First, it should be noted that in CALAN the threshold of channel n is called channel n-0.5. Thus, channel 100.0 in the CALAN output signifies the midpoint of channel 100, while channel 99.5 signifies the channel 100 threshold. The gain is defined as

$$G \equiv \text{gain} \equiv \frac{E(\text{channel } 999.5) - E(\text{channel } 99.5)}{(999.5 - 99.5)} \quad (2)$$

where E is the energy corresponding to channels 999.5 and 99.5 (CALAN channel definitions). To determine the offset and linearity, a line is drawn through the thresholds of channels 100 and 1000 on a channel vs. energy graph (see figure 3). The equation of this line is

$$C = \frac{1}{G} (E - E(999.5)) + 999.5 \quad (3)$$

where C is channel, E is energy of channel C, G is gain, and E(999.5) is the energy of the channel 1000 threshold. The offset is defined as the intercept (in channels) of this line with the channel axis. In terms of equation 3, the offset can be expressed as

$$0 = \text{Offset} = 999.5 - \frac{1}{G} E(999.5).$$




Figure 3

The linearity is defined as the difference (in channels) between channel 19.5 (=channel 20 threshold) and the channel predicted from equation 3 corresponding to the measured channel 20 threshold energy. In terms of equation 3 the linearity is

$$L = 19.5 - \frac{1}{G} [E(19.5) - E(999.5)] + 999.5$$

$$= - \frac{1}{G} [E(19.5) - E(999.5)] - 980$$

III. Results of Electronic Calibrations

Appendix A-H (A-J) contains a sample of the IMP-H (IMP-J) CALAN output for selected detectors. The complete set of all IMP-H (IMP-J) CALAN output is stored in a large black (blue) loose leaf notebook. Two calibrations are included for D2 and D5 because of instabilities in the calibration of these detectors (see discussion below). Table #1 lists preflight electronic calibrations performed on IMP-H, and Table #2 gives the corresponding data for IMP-J. Other information in Tables #1 and #2 is discussed below.

The thresholds of channels not calibrated directly can be obtained by linear interpolation between measured thresholds. Gordon Hurford's program ABRET performs this task. Appendix B-H contains the measured and interpolated thresholds for channels 1-100 for all IMP-H detectors. The corresponding data for IMP-J is in Appendix B-J. The CALAN output used for each detector is listed in Tables #1 and #2. The IMP-H D2 and D5 thresholds are based on 2/18/72 CALAN data with a shift of -9 and +10 keV per channel respectively. The thresholds in Appendices A-H and A-J can be used for interpreting flight data (see D2, D5 exceptions below). If highly accurate threshold values for prelaunch

Table 1 - IMP-H ELECTRONIC CALIBRATIONS

<u>DATE</u>	<u>D2 OFFSET (channels)</u>	<u>D2 GAIN (KeV/ch)</u>	<u>D5 OFFSET (channels)</u>	<u>D5 GAIN (KeV/ch)</u>	<u>Source of data for Appendix B-H</u>
6/29/71	6.274	41.146	2.374	41.358	
7/13/71	6.121	41.148	2.023	41.352	
7/19/71	6.205	41.156	4.187	41.367	D \emptyset , D8
2/18/72	6.149	41.148	1.438	41.402	D2*, D5*
2/23/72	12.438	41.225	1.635	41.399	
2/28/72	10.600	41.140	4.354	41.405	
3/2/72	11.887	41.153	4.527	41.408	D3
3/13/72	10.624	41.161	4.249	41.434	D1, D4, D6, D7, D9

*Corrected (see text)

Table 2 - IMP-J ELECTRONIC CALIBRATIONS

<u>DATE</u>	<u>D2 OFFSET (channels)</u>	<u>D2 GAIN (KeV/ch)</u>	<u>D5 OFFSET (channels)</u>	<u>D5 GAIN (channels)</u>	<u>Source of data for Appendix B-J</u>
10/31/72	8.255	40.999	0.887	41.256	
11/6/72	8.250	41.054	0.900	41.319	
11/19/72	8.354	41.011	0.977	41.286	
11/27/72	8.261	41.055	0.844	41.372	
12/11/72	8.330	41.056	0.919	41.366	
12/15/72	8.338	41.028	0.956	41.309	
1/2/73	6.998	41.032	-0.281	41.334	
6/26/73	8.320	40.971	0.956	41.194	All Detectors

data are required the CALAN output for the appropriate period should be consulted (see Tables #1 and #2). However, except for D2 and D5, CALAN threshold values were very stable with time.

IV. The Calibration of D2 and D5

Both the pre-flight calibrations and post-launch data contain examples of sudden shifts in the threshold of the individual channels of D2 and D5. (the only detectors with 16 MHz ADC's). These calibration shifts appear to involve a sudden shift in the ADC offset by $\pm \lesssim 3$ channels, while the ADC gain remains essentially unchanged. (See, for example, the IMP-H CALAN output of 2/23/72 and 2/28/72 in Appendix A-H.) The cause of this problem is not understood. The few examples available indicate that it is possible to use the unstimulated output of the detector involved to monitor the occurrence and magnitude of these shifts. The D5 unstimulated output (D5U) is read out during PLO and NEUTRAL events, while the D2 unstimulated output for IMP-H (D2U) is read out during D5 single events. Figures 4 and 5 show D5U for IMP-H as a function of time, and figures 6 and 7 show D2U as a function of time. This data is also summarized in Table 3. Note that D2U has undergone at least 2 major shifts since the launch of IMP-H.


We consider now D5 on IMP-H in some detail. Figure 8 shows a cross-plot of the CALAN D5 offset (D50) vs the mean D5U channel # obtained from either PARET or particle calibration (β, γ) data taken on the same or a nearby day. The few examples available suggest that there is a close correspondence between D50 and D5U.

Table 3. - D2 and D5 Unstimulated Pulse Heights - IMP-H

<u>Dates</u>	<u>Source of Data</u>	Mean D2 <u>Unstim.</u> <u>Ch.#</u>	Mean D5 <u>Unstim.</u> <u>Ch.#</u>	<u>Comments</u>
6/27	PARET	6.60	-	
6/29/71	PARET	6.92	1.93	
71:293	PARET	6.24		
:355	PARET	6.45		
1/30/72	γ -DATA		.29	
2/18	PARET	6.36	1.02	
2/20	γ		.97	
2/26	β		<.4	
2/27-28	β		4.34	
2/28	PARET	10.6,11.12	4.39	Unstable
3/2	PARET	11.91	3.90	
3/2-3	γ		3.93	
3/6	γ		3.98	
3/6-8	β		4.00	
3/9-10	β		3.94	
3/12-13	γ		3.91	
3/13	PARET	10.72	3.97	
4/21/72	PARET	11.98		
4/24/72	PARET	10.15		
4/25	β		4-6	D5 Double Valued at Times
4/26	β		4.15	
Launch 9/25/72				
72:273	Flight Data		.9	
293		6.76	.90	
303-12		7.05		
329-32		6.94		
333-36		7.13		
351-52		7.06		
73:007		6.88		
16-19		6.85	.90	
20-21		6.85	.91	
45-46		6.91	.84	
48-53		6.84	.97	
71-73		6.88		
78-81		7.04	.92	
89-91		10.79		
101-112		10.86	.90	Note: No change in D5U when D2U changes
114-116		10.49		
119-120		10.32		
137-141		10.40		
73:141-144		10.36		
161-163		7.46		
164-168		7.38		
171		7.44		


Note: A second shift of D2U occurs at about 75:100.

1972 - IMP-H



10 X 10 TO 1/2 INCH 46 1322
TELETYPE ASSEMBLY

1973 - IMP-H



DAY OF YEAR →

0 20 40 60 80 100 120 140 160 180 200 220 240 260 280 300 320 340 360

1973

It has been found empirically that it is possible to a good approximation to map preflight CALAN channel thresholds taken on different dates onto each other by simply shifting the calibrations by the # of channels that D5U has shifted. This procedure is necessary to obtain corrected channel thresholds for flight time periods when there is evidence of a drift in D5U. The recommended procedure is to select the CALAN output corresponding to a D5U value similar to that on the day in question. For example, on 2/18/73 we have $D5U_0 = 1.02$ channels (PARET output, Table 3) and the D5 CALAN output in Appendix A-H. To obtain individual channel calibrations for day x when $D5U = D5U_x$ we define

$$\Delta U_x = D5U_x - 1.02$$

and the threshold $E_x(n)$ of channel n on day x is given by $E_x(n) = E_0(n - \Delta U_x)$ where $E_0(n - \Delta U_x)$ is the interpolated threshold of channel $n - \Delta U_x$ in the 2/18/73 CALAN output. Because of the nonlinearity of low D5 channels, channel by channel interpolation is necessary. This simple algorithm leads to errors of $\lesssim 10$ KeV for channels #'s $n < 100$ in all cases where direct comparison with other CALAN output is possible. For $n > 100$ the error is $< .3\%$. This level of accuracy should be sufficient for almost all uses.

Application of this procedure to IMP H D5 electron pulse height spectra from β -spectrometer data taken during periods of different D5U levels ($\Delta U = 4$ channels) gave consistent measurements of the beam energy to within 20 KeV for .25 and .50 MeV electrons. Uncertainties in the β -spectrometer magnet calibration are of this order.

FIGURE 10 X 10 TO 1/2 INCH 46 1322
KCIUFEL & ESSER INC.

DS OFFSET




FIGURE 10 X 10 TO 1/2 INCH 46 1322
KCIUFEL & ESSER INC.

D2 OFFSET




Figure 9 shows a comparison of the D2U and D20 values measured on the same days. The discrepancies are of the order of a few tenths of a channel. From Table 3 and Figures 6 and 7 it is apparent that D2U is typically in either the range 6-8 or 10-12. D2 channel thresholds accurate to < 10 KeV can be obtained for any time period using the interpolation procedure described for D5. For D2U ~ 7 the 2/18/73 CALAN output (D2U = 6.36) should be used, while for D2U ~ 11 the 3/13/72 CALAN output (D2U = 10.72) would be appropriate.

The numbers in Table 3 and Figures 4 to 7 should be used only as a guide to offset shifts that have been observed. If accurate thresholds are desired when working with D2 or D5 data from either instrument, the D2U and D5U levels within that data set should be determined and the appropriate corrections made on as fine a time scale as is necessary.

With IMPJ the D2U value is lost. Therefore, some other technique, such as end of range proton channel #'s would have to be used to determine if any calibration shift occurs in post-launch data.

V. Problems in Interpreting Data

In this section three problems related to the use of calibrations in interpreting real data will be discussed. They are 1) overflows, 2) underflows, and 3) rundown not reaching an accumulator.

An overflow occurs when the energy loss in a detector saturates the pulse height analyzer (PHA). For all detectors except D2 and D5, when this happens the final state is either channel 1022 or 1023 (scale of 0 to 1023). The final states for D2 and D5 are 3 and 4092 respectively (scale of 0 to 1023 and 0 to 4092). (Note that D2's high channels are ...1022, 1023, 0, 1, 2, 3.) Since the D2H

threshold is channel 16, and the D2 offset for a typical calibration is + 10 channels, the first 3 channels of D2 correspond to overflows only, and not to small pulse heights.

An underflow occurs when the ADC rundown signal is shorter than the nominally allowed value. The reason for this malfunction isn't understood although it may be a pulse shape or timing problem. All accumulators are preset to some small negative value (equivalent to a large positive value in the binary code) (see tables 4 and 5). The minimum rundown should be long enough for the PHA to count to channel 0. When a shorter rundown is accepted, the PHA is in a large positive (or equivalently, a small negative) channel. The most common channel is -1 for all detectors except D2. This is read out as (is equivalent to) channel 4092 for D5 (D5 channels are multiples of 4 above 1024). A priori such an event is indistinguishable from an overflow. Although channel -1 is the most common underflow channel, other values have been observed (see tables 4 and 5).

Occasionally, for some unknown reason, a rundown signal doesn't reach the appropriate accumulator. In this case the channel in the PHA corresponds to the preset value in the accumulator. For some even less understood reason, some spurious counts are often added to this preset value. For example, the D6 accumulator is preset to channel 1011, but according to the phenomenon described above channel 1012 may be read out. Table 4 lists the observed occurrences of this phenomenon.

Table 4. - IMP-H Underflows and Overflows

<u>Detector</u>	<u>Negative Preset Value</u>	<u>Unstim. Ch.#.</u>	<u>Observed Preset Ch.#'s</u>	<u>Observed Underflows</u>	<u>Range Possibly Affected by Underflows</u>	<u>Overflow Ch.#'s</u>
D \emptyset	1012(-12)	0,1 [*]	1012	1020 - 1023	1013 - 1023	1022,1023
D2	979(-45)	See Fig. 4-5	979		980 - 1023	3 ⁺
D5 [†]	4048(-48)	See Fig. 6-7	4048	4064,4068, 4088, 4092	4052 - 4092 in multiples of 4	4092
D7	1011(-13)	0	1011	1012 - 1014 1020 - 1023	1012 - 1023	1022,1023
All others	1011(-13)	0			1012 - 1023	1022,1023

NOTES:

* Channel \emptyset equivalent to channel 1 in detector D \emptyset

[†] For D2 only, interpret channel 0 as 1024

1 as 1025

2 as 1026

3 as > 1026

[‡] Channel \emptyset in D5H is equivalent to an overflow

Miscellaneous

For lack of better place to put it, some information pertaining to channel calibrations is given in this section.

D2 and D5 have a second threshold called D2H and D5H (H stands for high), which are important for the priority and rate equations. The D2H threshold is channel 16. The D6H threshold, as determined from calibration data, is approximately channel 80. Because of timing uncertainties, the D5H threshold can vary by several channels.

Except for errors, such as underflows, the PHA channel for each detector should be at least as great as a certain minimum value. Tables 4 and 5 list the channel numbers that are read out for detectors that have not been triggered. Note that on IMPJ, D0 is never channel 0, and for D5 is seldom channel 0. (See internal report #50 for detailed explanation.) The IMPJ unstimulated value for D2 is unknown since the D2 channel is read out only if D2H triggers.

Table 5. - IMPJ Underflows and Overflows

<u>Detector</u>	<u>Preset Value</u>	<u>Unstim. Ch.#</u>	<u>Observed Preset Ch.#'s</u>	<u>Observed Underflows</u>	<u>Range Possibly Affected by Underflows</u>	<u>Overflow Ch.#</u>
D0	1013(-11)	1,2			1014 - 1023	1022,1023
D2	979(-45)		979		980-1023	3
D5*	4048(-48)	0-4	4048	4056,4092	4052-4092 in multiples of 4	4092
D7	1011(-13)	0			1012-1023	1022,1023
all others	1011(-13)	0		1013	1012-7023	1022,1023

*Channel 0 in D5H is equivalent to an overflow.

APPENDIX A-H

SELECTED IMP-H CALAN OUTPUT

This appendix contains selected CALAN output for IMP-H detectors, including gain offset, and linearity values, and calibrated channel thresholds in MeV. The ADC threshold values should be disregarded (see Section I-B). Prelaunch calibration dates are summarized in Table 1. Two calibrations are included for D2 and D5 as discussed in Section IV. Thresholds for intermediate channels can be found in Appendix B-H.

H DETECTOR #00

7-19-71 POST THERMAL VACUUM CALIBRATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01563 VOLTS 158.5 KEV

SIGMA = 0.00229 VOLTS 22.4 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
2	0.03370	0.3257
3	0.04716	0.4574
4	0.06320	0.6145
5	0.07966	0.7756
6	0.09612	0.9368
7	0.11292	1.1013
8	0.12965	1.2651
9	0.14659	1.4309
10	0.16365	1.5980
20	0.33242	3.2503
30	0.50130	4.9038
40	0.66949	6.5505
50	0.83798	8.2001
60	1.00620	9.8471
70	1.17408	11.4908
80	1.34234	13.1382
90	1.51020	14.7816
100	1.67834	16.4278
200	3.35977	32.8902
300	5.04132	49.3538
400	6.72220	65.8108
500	8.40334	82.2703
600	10.08421	98.7273
700	11.76451	115.1786
800	13.44516	131.6333
900	15.12653	148.0951
1000	16.80903	164.5680

GAIN 164.600 KEV/CHANNEL

OFFSET -0.305 CHANNELS (ENERGY)

LINEARITY +0.058 CHANNELS (ENERGY)

H DETECTOR #01

3-13-72 NEW D1 INSTALLED

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01691 VOLTS 160.8 KEV

SIGMA = 0.00240 VOLTS 22.9 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03160	0.3075
2	0.04705	0.4552
3	0.06390	0.6163
4	0.08071	0.7770
5	0.09808	0.9430
6	0.11523	1.1070
7	0.13281	1.2750
8	0.14980	1.4375
9	0.16715	1.6033
19	0.34086	3.2639
29	0.51420	4.9209
39	0.68716	6.5743
49	0.85979	8.2245
59	1.03237	9.8743
69	1.20515	11.5259
79	1.37777	13.1761
89	1.55024	14.8248
99	1.72308	16.4770
199	3.45029	32.9880
299	5.17824	49.5062
399	6.90627	66.0250
499	8.63413	82.5423
599	10.36233	99.0628
699	12.09088	115.5866
799	13.81997	132.1156
899	15.55087	148.6619
999	17.28332	165.2270

GAIN 165.273 KEV/CHANNEL

OFFSET -1.196 CHANNELS (ENERGY)

LINEARITY -0.053 CHANNELS (ENERGY)

--
H DETECTOR #02

18 FEB 1972 D7 DISCONNECTED; D3 DISABLED FOR ALL BUT D3

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01797 VOLTS 158.5 KEV

SIGMA = 0.00259 VOLTS 24.5 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
9	0.02211	0.1932
10	0.02392	0.2103
11	0.02615	0.2314
12	0.02920	0.2602
13	0.03248	0.2912
23	0.07235	0.6682
33	0.11535	1.0748
43	0.15897	1.4873
53	0.20240	1.8980
63	0.24640	2.3141
73	0.28984	2.7248
83	0.33350	3.1377
93	0.37727	3.5516
103	0.42096	3.9647
203	0.85626	8.0809
303	1.29131	12.1947
403	1.72629	16.3079
503	2.16144	20.4227
603	2.59658	24.5373
703	3.03160	28.6509
803	3.46683	32.7664
903	3.90220	36.8833
1003	4.33738	40.9983

GAIN 41.148 KEV/CHANNEL

OFFSET +6.149 CHANNELS (ENERGY)

LINEARITY +0.111 CHANNELS (ENERGY)

H DETECTOR #02

3-13-72 NEW DI INSTALLED

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01743 VOLTS 153.4 KEV

SIGMA = 0.00275 VOLTS 26.0 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
15	0.02400	0.2110
16	0.02702	0.2396
17	0.03024	0.2700
18	0.03391	0.3048
19	0.03743	0.3380
20	0.04139	0.3755
21	0.04551	0.4144
22	0.04964	0.4535
23	0.05377	0.4925
33	0.09642	0.8958
43	0.13978	1.3059
53	0.18325	1.7169
63	0.22712	2.1317
73	0.27057	2.5426
83	0.31436	2.9567
93	0.35817	3.3710

GAIN INSUFFICIENT DATA

OFFSET INSUFFICIENT DATA

LINEARITY INSUFFICIENT DATA

ER 07D7

H DETECTOR #03

3-2-72 POST VIBRATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01795 VOLTS 165.1 KEV

SIGMA = 0.00237 VOLTS 22.9 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03484	0.3163
2	0.04733	0.4368
3	0.06211	0.5795
4	0.07833	0.7360
5	0.09520	0.8987
6	0.11224	1.0631
7	0.12923	1.2271
8	0.14649	1.3936
9	0.16363	1.5590
19	0.33529	3.2153
29	0.50686	4.8707
39	0.67788	6.5208
49	0.84897	8.1715
59	1.01969	9.8187
69	1.19047	11.4665
79	1.36134	13.1152
89	1.53203	14.7621
99	1.70273	16.4091
199	3.41099	32.8914
299	5.11824	49.3640
399	6.82456	65.8275
499	8.52945	82.2773
599	10.23300	98.7142
699	11.93517	115.1377
799	13.63701	131.5580
899	15.33844	147.9744
999	17.03930	164.3854

GAIN 164.418 KEV/CHANNEL

OFFSET -1.301 CHANNELS (ENERGY)

LINEARITY +0.246 CHANNELS (ENERGY)

H DETECTOR #04

3-13-72 NEW DI INSTALLED

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01567 VOLTS 163.0 KEV

SIGMA = 0.00234 VOLTS 22.3 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03401	0.3329
2	0.04985	0.4842
3	0.06665	0.6446
4	0.08404	0.8107
5	0.10149	0.9773
6	0.11931	1.1475
7	0.13665	1.3131
8	0.15447	1.4833
9	0.17206	1.6513
19	0.34853	3.3366
29	0.52491	5.0211
39	0.70068	6.6997
49	0.87647	8.3786
59	1.05171	10.0521
69	1.22759	11.7318
79	1.40303	13.4073
89	1.57858	15.0838
99	1.75415	16.7606
199	3.51118	33.5405
299	5.26908	50.3288
399	7.02699	67.1172
499	8.78465	83.9032
599	10.54177	100.6840
699	12.29756	117.4521
799	14.05303	134.2172
899	15.80759	150.9736
999	17.56048	167.7140

GAIN 167.726 KEV/CHANNEL

OFFSET -1.428 CHANNELS (ENERGY)

LINEARITY +0.035 CHANNELS (ENERGY)

H DETECTOR #05

18 FEB 1972 D7 DISCONNECTED

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01597 VOLTS 159.5 KEV

SIGMA = 0.00166 VOLTS 15.6 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
7	0.02606	0.2582
8	0.02930	0.2886
9	0.03285	0.3219
10	0.03663	0.3574
20	0.07831	0.7483
30	0.12215	1.1595
40	0.16648	1.5753
50	0.21060	1.9892
60	0.25501	2.4057
70	0.29947	2.8227
80	0.34373	3.2379
90	0.38791	3.6523
100	0.43225	4.0682
200	0.87381	8.2098
300	1.31484	12.3465
400	1.75559	16.4806
500	2.19701	20.6210
600	2.63828	24.7599
700	3.07956	28.8990
800	3.52062	33.0360
900	3.96185	37.1746
1000	4.40317	41.3140
2000	8.81626	82.7072
3000	13.22929	124.0999
4000	17.64615	165.5284

GAIN 41.402 KEV/CHANNEL

OFFSET +1.438 CHANNELS (ENERGY)

LINEARITY -0.197 CHANNELS (ENERGY)

H. DETECTOR #05

3-13-72 NEW D1 INSTALLED

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01673 VOLTS 166.7 KEV

SIGMA = 0.00168 VOLTS 15.8 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
8	0.02069	0.2079
9	0.02324	0.2318
10	0.02625	0.2600
11	0.02950	0.2905
12	0.03324	0.3256
13	0.03708	0.3616
14	0.04107	0.3990
15	0.04519	0.4377
16	0.04934	0.4766
17	0.05349	0.5155
18	0.05781	0.5560
19	0.06197	0.5951
20	0.06631	0.6358
30	0.10994	1.0450
40	0.15412	1.4594
50	0.19843	1.8750
60	0.24268	2.2901
70	0.28707	2.7064
80	0.33140	3.1222
90	0.37542	3.5351
100	0.41978	3.9512
200	0.86205	8.0995
300	1.30335	12.2388
400	1.74451	16.3767
500	2.18615	20.5191
600	2.62761	24.6599
1000	4.39403	41.2283
2000	8.81050	82.6532
4000	17.64714	165.5377

GAIN 41.434 KEV/CHANNEL

OFFSET +4.249 CHANNELS (ENERGY)

LINEARITY -0.111 CHANNELS (ENERGY)

H DETECTOR #06

3-13-72 NEW DI INSTALLED

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01856 VOLTS 164.2 KEV

SIGMA = 0.00216 VOLTS 20.4 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03877	0.3467
2	0.05224	0.4740
3	0.06786	0.6217
4	0.08476	0.7814
5	0.10186	0.9431
6	0.11948	1.1096
7	0.13680	1.2733
8	0.15451	1.4407
9	0.17188	1.6049
19	0.34752	3.2652
29	0.52260	4.9202
39	0.69734	6.5719
49	0.87204	8.2233
59	1.04638	9.8713
69	1.22086	11.5206
79	1.39531	13.1696
89	1.56993	14.8202
99	1.74447	16.4701
199	3.49027	32.9725
299	5.23695	49.4833
399	6.98392	65.9968
499	8.73145	82.5155
599	10.47894	99.0339
699	12.22726	115.5602
799	13.97652	132.0953
899	15.72775	148.6491
999	17.48046	165.2168

GAIN 165.274 KEV/CHANNEL

OFFSET -1.153 CHANNELS (ENERGY)

LINEARITY -0.103 CHANNELS (ENERGY)

H DETECTOR #07

3-13-72 NEW D1 INSTALLED

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01235 VOLTS 143.6 KEV

SIGMA = 0.00188 VOLTS 18.0 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.01911	0.2063
2	0.03362	0.3454
3	0.04926	0.4954
4	0.06616	0.6575
5	0.08321	0.8210
6	0.10033	0.9852
7	0.11738	1.1487
8	0.13460	1.3139
9	0.15206	1.4813
19	0.32524	3.1422
29	0.49795	4.7986
39	0.67052	6.4536
49	0.84287	8.1065
59	1.01515	9.7588
69	1.18720	11.4088
79	1.35943	13.0606
89	1.53162	14.7120
99	1.70365	16.3618
199	3.42576	32.8776
299	5.14781	49.3929
399	6.86890	65.8990
499	8.58901	82.3956
599	10.30837	98.8851
699	12.02692	115.3668
799	13.74534	131.8472
899	15.46438	148.3305
999	17.18265	164.8127

GAIN 164.945 KEV/CHANNEL

OFFSET -0.695 CHANNELS (ENERGY)

LINEARITY +0.145 CHANNELS (ENERGY)

H DETECTOR #08

7-13-71 POST COLD, SOAK D3 DISABLED

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01713 VOLTS 138.8 KEV

SIGMA = 0.00237 VOLTS 22.4 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03804	0.3526
2	0.05070	0.4723
3	0.06549	0.6121
4	0.08208	0.7690
5	0.09917	0.9306
6	0.11663	1.0958
7	0.13416	1.2615
8	0.15190	1.4293
9	0.16913	1.5922
19	0.34499	3.2552
29	0.52016	4.9117
39	0.69538	6.5687
49	0.87006	8.2206
59	1.04480	9.8730
69	1.21949	11.5249
79	1.39405	13.1757
89	1.56877	14.8279
99	1.74342	16.4795
199	3.49082	33.0037
299	5.23873	49.5328
399	6.98658	66.0613
499	8.73420	82.5877
599	10.48224	99.1180
699	12.23006	115.6463
799	13.97813	132.1769
899	15.72703	148.7153
999	17.47752	165.2688

GAIN 165.322 KEV/CHANNEL

OFFSET -1.181 CHANNELS (ENERGY)

LINEARITY -0.009 CHANNELS (ENERGY)

H DETECTOR #09

3-13-72 NEW D1 INSTALLED

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01205 VOLTS 97.2 KEV

SIGMA = 0.00218 VOLTS 20.8 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03342	0.3029
2	0.04689	0.4315
3	0.06227	0.5784
4	0.07880	0.7363
5	0.09585	0.8991
6	0.11310	1.0638
7	0.13047	1.2297
8	0.14779	1.3951
9	0.16514	1.5608
19	0.33889	3.2201
29	0.51246	4.8777
39	0.68571	6.5322
49	0.85862	8.1835
59	1.03166	9.8360
69	1.20426	11.4844
79	1.37711	13.1351
89	1.54990	14.7852
99	1.72271	16.4355
199	3.45116	32.9422
299	5.17973	49.4499
399	6.90797	65.9545
499	8.63591	82.4563
599	10.36344	98.9541
699	12.09151	115.4571
799	13.82013	131.9654
899	15.55027	148.4882
999	17.28111	165.0176

GAIN 165.091 KEV/CHANNEL

OFFSET -1.054 CHANNELS (ENERGY)

LINEARITY +0.049 CHANNELS (ENERGY)

APPENDIX A-J

SELECTED IMP-J CALAN OUTPUT

This appendix contains selected CALAN output for IMP-J detectors, including gain offset, and linearity values, and calibrated channel thresholds in MeV. The ADC threshold values should be disregarded (see Section I-B). Prelaunch calibration dates are summarized in Table 1. Thresholds for intermediate channels can be found in Appendix B-J.

J DETECTOR #00

6-26-73 POST - DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01781 VOLTS 172.3 KEV

SIGMA = 0.00164 VOLTS 15.5 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
3	0.02960	0.2846
4	0.04534	0.4330
5	0.06212	0.5912
6	0.07946	0.7546
7	0.09683	0.9184
9	0.13158	1.2460
20	0.32435	3.0635
100	1.72171	16.2377
1000	17.42906	164.3260

GAIN 164.542 KEV/CHANNEL

OFFSET + 0.816 CHANNELS (ENERGY)

LINEARITY +0.016 CHANNELS (ENERGY)

J DETECTOR #01

6-26-73 POST - DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01768 VOLTS 168.0 KEV

SIGMA = 0.00165 VOLTS 15.2 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03304	0.3145
2	0.04962	0.4674
3	0.06728	0.6302
4	0.08505	0.7941
5	0.10310	0.9606
6	0.12116	1.1271
7	0.13911	1.2926
8	0.15740	1.4613
9	0.17554	1.6286
20	0.37501	3.4680
100	1.82101	16.8026
1000	18.05638	166.5203

GAIN 166.353 KEV/CHANNEL

OFFSET - 1.506 CHANNELS (ENERGY)

LINEARITY +0.158 CHANNELS (ENERGY)

J. DETECTOR #02

6-26-73 POST DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01498 VOLTS 162.7 KEV

SIGMA = 0.00264 VOLTS 24.8 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
17	0.03281	0.3443
18	0.03675	0.3813
19	0.04082	0.4196
20	0.04494	0.4583
21	0.04915	0.4979
22	0.05332	0.5371
23	0.05746	0.5760
25	0.06607	0.6569
100	0.39364	3.7357
1000	4.31681	40.6098

GAIN 40.971 KEV/CHANNEL

OFFSET + 8.320 CHANNELS (ENERGY)

LINEARITY +0.122 CHANNELS (ENERGY)

J DETECTOR #03

6-26-73 POST - DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01767 VOLTS 172.3 KEV

SIGMA = 0.00159 VOLTS 15.0 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03020	0.2910
2	0.04612	0.4407
3	0.06330	0.6023
4	0.08057	0.7647
5	0.09813	0.9299
6	0.11573	1.0954
7	0.13350	1.2625
8	0.15122	1.4292
9	0.16887	1.5952
20	0.36392	3.4295
100	1.77711	16.7200
1000	17.68728	166.3486

GAIN 166.254 KEV/CHANNEL

OFFSET - 1.069 CHANNELS (ENERGY)

LINEARITY -0.059 CHANNELS (ENERGY)

J DETECTOR #04

6-26-73 POST - DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01641 VOLTS 172.3 KEV

SIGMA = 0.00151 VOLTS 14.0 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (KEV)
1	0.03180	0.3210
2	0.04816	0.4731
3	0.06529	0.6323
4	0.08292	0.7962
5	0.10044	0.9591
6	0.11826	1.1247
7	0.13529	1.2908
8	0.15145	1.4541
9	0.17145	1.6198
10	0.36789	3.4452
100	1.79278	16.6908
1000	17.90143	166.4338

GAIN 166.381 KEV/CHANNEL

OFFSET - 0.816 CHANNELS (ENERGY)

LINEARITY -0.391 CHANNELS (ENERGY)

J DETECTOR #05

6-26-73 POST DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01790 VOLTS 193.4 KEV

SIGMA = 0.00191 VOLTS 17.9 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
5	0.02135	0.2219
6	0.02344	0.2415
7	0.02648	0.2700
8	0.02997	0.3028
9	0.03370	0.3378
10	0.03759	0.3742
11	0.04151	0.4110
12	0.04564	0.4497
13	0.04975	0.4883
14	0.05402	0.5283
15	0.05817	0.5673
20	0.07966	0.7688
100	0.43175	4.0709
400	1.74822	16.4177
4000	17.56061	164.7166

GAIN 41.194 KEV/CHANNEL

OFFSET + 0.956 CHANNELS (ENERGY)

LINEARITY -0.279 CHANNELS (ENERGY)

J DETECTOR #06

6-26-73 POST - DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01742 VOLTS 173.5 KEV

SIGMA = 0.00162 VOLTS 15.5 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.02976	0.2887
2	0.04541	0.4386
3	0.06198	0.5974
4	0.07891	0.7595
5	0.09594	0.9227
6	0.11326	1.0886
7	0.13048	1.2536
8	0.14771	1.4186
9	0.16500	1.5843
20	0.35527	3.4070
100	1.73349	16.6102
1000	17.23298	165.0928

GAIN 164.981 KEV/CHANNEL

OFFSET - 1.179 CHANNELS (ENERGY)

LINEARITY +0.028 CHANNELS (ENERGY)

J DETECTOR #07

6-26-73 POST - DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01676 VOLTS 183.3 KEV

SIGMA = 0.00164 VOLTS 15.7 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03138	0.3221
2	0.04716	0.4730
3	0.06372	0.6313
4	0.08056	0.7923
5	0.09784	0.9575
6	0.11498	1.1213
7	0.13219	1.2859
8	0.14937	1.4501
9	0.16660	1.6148
20	0.35677	3.4329
100	1.73556	16.6146
1000	17.25227	164.9591

GAIN 164.827 KEV/CHANNEL

OFFSET - 1.300 CHANNELS (ENERGY)

LINEARITY -0.027 CHANNELS (ENERGY)

J DETECTOR #08

6-26-73 POST - DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01728 VOLTS 159.1 KEV

SIGMA = 0.00171 VOLTS 16.0 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03339	0.3082
2	0.04980	0.4620
3	0.06701	0.6232
4	0.08435	0.7857
5	0.10199	0.9510
6	0.11952	1.1152
7	0.13747	1.2834
8	0.15512	1.4488
9	0.17299	1.6162
20	0.36852	3.4482
100	1.78293	16.7004
1000	17.69723	165.8081

GAIN 165.675 KEV/CHANNEL

OFFSET - 1.302 CHANNELS (ENERGY)

LINEARITY -0.011 CHANNELS (ENERGY)

DETECTOR #09

6-26-73 POST DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

ADC THRESHOLD = 0.01596 VOLTS 172.3 KEV

SIGMA = 0.00169 VOLTS 16.2 KEV

CHANNEL #	THRESHOLD (VOLTS)	THRESHOLD (MEV)
1	0.03091	0.3161
2	0.04680	0.4682
3	0.06348	0.6279
4	0.08039	0.7897
5	0.09751	0.9536
6	0.11493	1.1204
7	0.13206	1.2843
8	0.14955	1.4517
9	0.16666	1.6155
20	0.35741	3.4415
100	1.73807	16.6576
1000	17.25339	165.1757

GAIN 165.020 KEV/CHANNEL

OFFSET - 1.443 CHANNELS (ENERGY)

LINEARITY +0.088 CHANNELS (ENERGY)

J DETECTOR #10

6-26-73 POST DETECTOR RECONFIGURATION

1000 COUNT CALIBRATION

DISC. THRESH. = 0.01399 VOLTS 132.6 KEV

SIGMA = 0.00129 VOLTS 12.2 KEV

APPENDIX B-H

IMP-H MEASURED AND INTERPOLATED CHANNEL THRESHOLDS

This appendix contains measured and interpolated channel thresholds obtained from ABRET (see Section III) for channels 1-114. These thresholds are appropriate for analyzing flight data. Notice that the thresholds for D2 and D5 have been shifted. (See Section IV.) Thresholds for channels >100 can be extrapolated from channel 100 using the gain and offset relationship (Section I).

IMP-H

CALIBRATION DATA FOR INDIVIDUAL ADC'S

CHAN	D0	D1	D2	D3	D4	D5(III)	D6	D7	D8	D9	D5(I OR II)
SHIFT	0.0	0.0	-0.009	0.0	0.0	+0.010	0.0	0.0	0.0	0.0	+0.010
EFSET	-0.032	0.115	-0.265	0.132	0.156	-0.049	0.108	0.032	0.117	0.091	-0.090
GAIN	0.164600	0.165273	0.041148	0.164418	0.167726	0.165608	0.165274	0.164945	0.165330	0.165091	0.041402
0	0.158	0.161	0.158	0.165	0.163	0.1600	0.164	0.144	0.140	0.160	0.160
1	0.158A	0.308	0.158A	0.316	0.333	0.1600	0.347	0.206	0.346	0.303	0.160A
2	0.326	0.455	0.158A	0.437	0.484	0.3320	0.474	0.345	0.468	0.432	0.160A
3	0.457	0.616	0.158A	0.580	0.645	0.4840	0.622	0.495	0.607	0.578	0.160A
4	0.614	0.777	0.158A	0.736	0.811	0.6410	0.781	0.658	0.764	0.736	0.160A
5	0.776	0.943	0.158A	0.899	0.977	0.7990	0.943	0.821	0.925	0.899	0.160A
6	0.937	1.107	0.158A	1.063	1.148	0.9640	1.110	0.985	1.090	1.064	0.160A
7	1.101	1.275	0.158A	1.227	1.313	1.1290	1.273	1.149	1.255	1.230	0.268
8	1.265	1.438	0.158A	1.394	1.483	1.2940	1.441	1.314	1.421	1.395	0.299
9	1.431	1.603	0.184	1.559	1.651	1.4600	1.605	1.481	1.588	1.561	0.332
10	1.599	1.7691	0.201	1.7251	1.8201	1.6260	1.7711	1.6471	1.7541	1.7271	0.367
11	1.7631	1.9351	0.222	1.8901	1.9881	1.7920	1.9371	1.8131	1.9201	1.8931	0.4061
12	1.9281	2.1011	0.251	2.0561	2.1571	1.9580	2.1031	1.9791	2.0861	2.0591	0.4451
13	2.0941	2.2671	0.282	2.2211	2.3251	2.1240	2.2691	2.1451	2.2521	2.2251	0.4841
14	2.2591	2.4331	0.3201	2.3871	2.4941	2.2910	2.4351	2.3111	2.4181	2.3901	0.5231
15	2.4241	2.6001	0.3571	2.5531	2.6631	2.4580	2.6011	2.4731	2.5851	2.5561	0.5621
16	2.5891	2.7661	0.3951	2.7181	2.8311	2.6240	2.7671	2.6441	2.7511	2.7221	0.6021
17	2.7541	2.9321	0.4331	2.8841	3.0001	2.7910	2.9331	2.8101	2.9171	2.8881	0.6411
18	2.9201	3.0981	0.4701	3.0491	3.1681	2.9570	3.0991	2.9761	3.0831	3.0541	0.6801
19	3.0851	3.264	0.5081	3.215	3.337	3.1230	3.265	3.142	3.249	3.220	0.7191
20	3.250	3.4301	0.5461	3.3811	3.5051	3.2890	3.4301	3.3081	3.4151	3.3861	0.758
21	3.4151	3.5951	0.5841	3.5461	3.6741	3.4550	3.5961	3.4731	3.5811	3.5521	0.7991
22	3.5811	3.7611	0.6211	3.7121	3.8421	3.6210	3.7611	3.6391	3.7461	3.7171	0.8401
23	3.7461	3.9271	0.659	3.8771	4.0111	3.7870	3.9271	3.8051	3.9121	3.8831	0.8821
24	3.9121	4.0921	0.7001	4.0431	4.1791	3.9530	4.0921	3.9701	4.0781	4.0491	0.9231
25	4.0771	4.2581	0.7401	4.2091	4.3471	4.1190	4.2581	4.1361	4.2441	4.2151	0.9641
26	4.2421	4.4241	0.7811	4.3741	4.5161	4.2850	4.4231	4.3021	4.4101	4.3811	1.0051
27	4.4081	4.5901	0.8221	4.5401	4.6841	4.4510	4.5891	4.4681	4.5751	4.5461	1.0461
28	4.5731	4.7551	0.8621	4.7051	4.8531	4.6160	4.7541	4.6331	4.7411	4.7121	1.0881
29	4.7391	4.921	0.9031	4.871	5.021	4.7820	4.920	4.799	4.907	4.878	1.1291
30	4.904	5.0861	0.9441	5.0361	5.1891	4.9480	5.0851	4.9641	5.0721	5.0431	1.170
31	5.0691	5.2521	0.9851	5.2011	5.3571	5.1130	5.2501	5.1301	5.2381	5.2091	1.2111
32	5.2331	5.4171	1.0251	5.3661	5.5251	5.2790	5.4161	5.2951	5.4031	5.3741	1.2531
33	5.3981	5.5821	1.066	5.5311	5.6931	5.4450	5.5811	5.4611	5.5691	5.5401	1.2941
34	5.5621	5.7471	1.1071	5.6961	5.8601	5.6110	5.7461	5.6261	5.7341	5.7051	1.3361
35	5.7271	5.9131	1.1481	5.8611	6.0281	5.7760	5.9111	5.7921	5.8991	5.8701	1.3771
36	5.8921	6.0781	1.1901	6.0261	6.1961	5.9420	6.0761	5.9571	6.0651	6.0361	1.4191
37	6.0561	6.2431	1.2311	6.1911	6.3641	6.1080	6.2421	6.1231	6.2301	6.2011	1.4601
38	6.2211	6.4091	1.2721	6.3561	6.5321	6.2730	6.4071	6.2881	6.3961	6.3671	1.5021
39	6.3851	6.574	1.3131	6.521	6.700	6.4390	6.572	6.454	6.561	6.532	1.5431
40	6.550	6.7391	1.3541	6.6861	6.8681	6.6050	6.7371	6.6191	6.7261	6.6971	1.585
41	6.7151	6.9041	1.3961	6.8511	7.0361	6.7700	6.9021	6.7841	6.8921	6.8621	1.6261
42	6.8801	7.0691	1.4371	7.0161	7.2041	6.9360	7.0671	6.9501	7.0571	7.0281	1.6681
43	7.0451	7.2341	1.478	7.1811	7.3721	7.1020	7.2321	7.1151	7.2221	7.1931	1.7091
44	7.2101	7.3991	1.5191	7.3461	7.5391	7.2670	7.3971	7.2801	7.3871	7.3581	1.7511
45	7.3751	7.5641	1.5601	7.5121	7.7071	7.4330	7.5631	7.4451	7.5531	7.5231	1.7921
46	7.5401	7.7291	1.6011	7.6771	7.8751	7.5990	7.7281	7.6101	7.7181	7.6881	1.8331
47	7.7051	7.8941	1.6421	7.8421	8.0431	7.7640	7.8931	7.7761	7.8831	7.8541	1.8751
48	7.8701	8.0591	1.6831	8.0071	8.2111	7.9300	8.0581	7.9411	8.0491	8.0191	1.9161
49	8.0351	8.224	1.7251	8.172	8.379	8.0560	8.223	8.106	8.214	8.184	1.9581
50	8.200	8.3891	1.7661	8.3371	8.5461	8.2610	8.3881	8.2711	8.3791	8.3491	1.999
51	8.3651	8.5541	1.8071	8.5011	8.7141	8.4270	8.5531	8.4371	8.5451	8.5141	2.0411
52	8.5291	8.7191	1.8481	8.6661	8.8811	8.5520	8.7171	8.6021	8.7101	8.6801	2.0821

APPENDIX B-J

IMP-J MEASURED AND INTERPOLATED CHANNEL THRESHOLDS

This appendix contains measured and interpolated channel thresholds obtained from ABRET (see Section III) for channels 1-114. These thresholds are appropriate for analyzing flight data. Thresholds for channels >100 can be extrapolated from channel 100 using the gain and offset relationship (Section I).

IMP-5

CALIBRATION DATA FOR INDIVIDUAL AEC'S

CHAN	D0	D1	D2	D3	D4	D5(III)	D6	D7	D8	D9	D5(I CR II)
SHIFT	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
EFSET	-0.217	0.167	-0.261	0.095	0.053	-0.019	0.112	0.132	0.133	0.156	-0.060
GAIN	0.164542	0.166353	0.164971	0.166254	0.166381	0.164776	0.164981	0.164827	0.165675	0.165020	0.041194
0	0.168	0.164	C.150	0.160	C.164	0.1630	C.159	0.166	0.157	0.161	C.163
1	0.168A	0.314	C.150A	0.291	C.321	0.2220	C.289	C.222	0.208	C.316	C.163A
2	0.168A	0.467	C.150A	0.441	C.473	0.3390	0.439	C.473	0.462	0.468	C.163A
3	0.265	0.530	C.150A	0.602	C.632	0.4880	C.597	C.631	0.623	0.628	0.163A
4	0.433	C.794	C.150A	0.765	C.796	0.6480	C.799	C.792	C.786	0.790	C.163A
5	0.591	C.961	C.150A	C.930	C.956	C.8100	C.923	0.957	C.951	0.954	C.222
6	0.755	1.127	C.150A	1.095	1.125	0.9750	1.085	1.121	1.115	1.120	0.241
7	0.919	1.293	C.150A	1.262	1.289	1.1400	1.254	1.286	1.283	1.284	C.270
8	1.082I	1.461	C.150A	1.429	1.454	1.3050	1.419	1.450	1.449	1.452	C.303
9	1.246	1.629	C.150A	1.595	1.619	1.4700	1.584	1.615	1.616	1.615	C.338
10	1.411I	1.796I	C.150A	1.762I	1.785I	1.6360	1.750I	1.780I	1.783I	1.781I	C.374
11	1.576I	1.933I	C.150A	1.929I	1.951I	1.8010	1.916I	1.945I	1.949I	1.947I	C.411
12	1.742I	2.130I	C.150A	2.095I	2.117I	1.9660	2.081I	2.111I	2.116I	2.113I	C.450
13	1.907I	2.297I	C.150A	2.262I	2.283I	2.1310	2.247I	2.276I	2.282I	2.275I	C.488
14	2.072I	2.465I	C.150A	2.429I	2.449I	2.2960	2.413I	2.441I	2.449I	2.445I	C.528
15	2.237I	2.632I	C.150A	2.596I	2.615I	2.4610	2.578I	2.606I	2.615I	2.611I	C.567
16	2.403I	2.799I	C.308	2.762I	2.781I	2.6260	2.744I	2.772I	2.782I	2.777I	C.608I
17	2.568I	2.966I	0.344	2.929I	2.947I	2.7910	2.910I	2.937I	2.949I	2.943I	C.648I
18	2.733I	3.134I	C.391	3.096I	3.113I	2.9560	3.076I	3.102I	3.115I	3.109I	C.688I
19	2.898I	3.301I	C.420	3.263I	3.279I	3.1220	3.241I	3.268I	3.282I	3.275I	C.728I
20	3.063	3.468	C.458	3.429	3.445	3.2870	3.407	3.433	3.448	3.441	C.769
21	3.228I	3.635I	0.498	3.596I	3.611I	3.4520	3.572I	3.598I	3.614I	3.607I	C.810I
22	3.393I	3.801I	C.537	3.762I	3.776I	3.6170	3.737I	3.762I	3.780I	3.772I	C.851I
23	3.558I	3.968I	C.576	3.928I	3.942I	3.7820	3.902I	3.927I	3.945I	3.937I	C.893I
24	3.722I	4.135I	C.616I	4.094I	4.107I	3.9470	4.067I	4.092I	4.111I	4.102I	C.934I
25	3.887I	4.301I	C.657	4.260I	4.273I	4.1120	4.232I	4.257I	4.276I	4.268I	C.975I
26	4.052I	4.468I	C.698I	4.426I	4.439I	4.2770	4.397I	4.422I	4.442I	4.433I	1.016I
27	4.216I	4.635I	C.739I	4.592I	4.604I	4.4410	4.562I	4.586I	4.608I	4.598I	1.058I
28	4.381I	4.801I	C.780I	4.759I	4.770I	4.6060	4.727I	4.751I	4.773I	4.763I	1.099I
29	4.546I	4.968I	C.821I	4.925I	4.935I	4.7710	4.892I	4.916I	4.939I	4.928I	1.140I
30	4.710I	5.135I	C.862I	5.091I	5.101I	4.9350	5.057I	5.081I	5.105I	5.094I	1.182I
31	4.875I	5.302I	C.903I	5.257I	5.266I	5.1000	5.222I	5.245I	5.270I	5.259I	1.223I
32	5.040I	5.468I	C.944I	5.423I	5.432I	5.2640	5.387I	5.410I	5.436I	5.424I	1.264I
33	5.204I	5.635I	C.985I	5.589I	5.598I	5.4290	5.553I	5.575I	5.602I	5.589I	1.305I
34	5.369I	5.802I	1.026I	5.755I	5.763I	5.5940	5.718I	5.740I	5.767I	5.754I	1.347I
35	5.534I	5.968I	1.067I	5.921I	5.929I	5.7580	5.883I	5.904I	5.933I	5.920I	1.388I
36	5.698I	6.135I	1.108I	6.088I	6.094I	5.9230	6.048I	6.069I	6.099I	6.085I	1.429I
37	5.863I	6.302I	1.150I	6.254I	6.260I	6.0880	6.213I	6.234I	6.264I	6.250I	1.470I
38	6.028I	6.468I	1.191I	6.420I	6.425I	6.2520	6.378I	6.399I	6.430I	6.415I	1.512I
39	6.192I	6.635I	1.232I	6.586I	6.591I	6.4170	6.543I	6.564I	6.598I	6.580I	1.553I
40	6.357I	6.802I	1.273I	6.752I	6.757I	6.5810	6.708I	6.728I	6.761I	6.746I	1.594I
41	6.522I	6.968I	1.314I	6.918I	6.922I	6.7460	6.873I	6.893I	6.927I	6.911I	1.636I
42	6.686I	7.135I	1.355I	7.084I	7.088I	6.9110	7.038I	7.058I	7.093I	7.076I	1.677I
43	6.851I	7.302I	1.396I	7.251I	7.253I	7.0750	7.203I	7.223I	7.258I	7.241I	1.718I
44	7.016I	7.468I	1.437I	7.417I	7.419I	7.2400	7.368I	7.397I	7.424I	7.406I	1.759I
45	7.180I	7.635I	1.478I	7.583I	7.584I	7.4050	7.533I	7.552I	7.590I	7.572I	1.801I
46	7.345I	7.802I	1.519I	7.749I	7.750I	7.5690	7.698I	7.717I	7.755I	7.737I	1.842I
47	7.510I	7.968I	1.560I	7.915I	7.916I	7.7340	7.863I	7.882I	7.921I	7.902I	1.883I
48	7.674I	8.135I	1.601I	8.081I	8.081I	7.8980	8.028I	8.046I	8.086I	8.067I	1.925I
49	7.839I	8.302I	1.642I	8.247I	8.247I	8.0630	8.193I	8.211I	8.252I	8.232I	1.966I
50	8.004I	8.468I	1.683I	8.413I	8.412I	8.2280	8.358I	8.376I	8.418I	8.398I	2.007I
51	8.168I	8.635I	1.724I	8.580I	8.578I	8.3920	8.523I	8.541I	8.583I	8.563I	2.048I
52	8.333I	8.802I	1.765I	8.746I	8.743I	8.5570	8.688I	8.706I	8.749I	8.728I	2.090I

